PROFESSIONAL MEASURING
SINCE 1844

KERN

Product brochure

31.10.2026

KERN KFP 300V30M

Platform: entirely out of stainless steel

ENGLISH



CONTACT

KERN & SOHN GmbH
Ziegelei 1
72336 Balingen
Germany

+49 7433 9933-0
info@kern-sohn.com
www.kern-sohn.com







FEATURES

- Platform: entirely out of stainless steel
- 1 weighing cell, stainless steel, silicone-coated, IP67, OIML R60 approval for verification, class III, 3000 e
- Level indicator and levelling feet for precise levelling of the scale, fitted as standard, to give the most accurate weighing result

TECHNICAL DATA

- Weighing platform dimensions W×D×H 650×500×139 mm, stainless steel
- Connection cable approx. 3 m
- Net weight approx. 18 kg

PICTOGRAMS

Standard



Factory



MODEL OVERVIEW

Model	Weighing capacity [Max]	Readability [d]	Verification value [e]	Minimal load [Min]	Cable length	Dimensions weighing surface	Net weight approx.
KFP 15V30M	15 kg	1 g	5 g	100 g	3 m	300×240 mm	8 kg
KFP 30V30M	30 kg	2 g	10 g	200 g	3 m	400×300 mm	8 kg
KFP 30V30SM	30 kg	2 g	10 g	200 g	3 m	300×240 mm	5,0 kg
KFP 60V30M	60 kg	5 g	20 g	400 g	3 m	400×300 mm	8 kg
KFP 150V30M	150 kg	10 g	50 g	1 kg	3 m	500×400 mm	10 kg
KFP 150V30SM	150 kg	10 g	50 g	1 kg	3 m	400×300 mm	10 kg
KFP 300V30M	300 kg	20 g	100 g	2 kg	3 m	650×500 mm	18 kg

ACCESSORIES

Model	Description
YGR-01	ESD grounding plug incl. grounding cable YGR-01
YKV-01	Digital weighing transmitter YKV-01
YKV-02	Digital weighing transmitter YKV-02
YRO-03	roller conveyor YRO-03

SERVICES

Model	Description
963-129	DAkkS calibration certificate 963-129
965-229	Verification / Conformity assessment according to NAWI (2014/31/EU) 965-229
965-403	Cable extension 965-403 (< 4 m)
965-404	Cable extension 965-404 (4–7 m)
965-405	Cable extension 965-405 (7–20 m)
965-412	Assembly of components 965-412 (> 50 kg - 350 kg)
970-106	Warranty extension (+2 years) 970-106

KERN Pictogram Overview

	Protection against dust and water splashes IP67
	Package shipment by courier service: The duration of internal product provision is one day
	Conformity assessment: The duration of the conformity assessment is 3 working days